

SIEMENS MAGNETOM Skyra syngo MR D13

\\USER\CIND\StudyProtocols\PTSA*dm_ep2d_mono95_b0_p2_iso2.0
TA:1:14 PAT:2 Voxel size:2.0x2.0x2.0 mm Rel. SNR:1.00 :epse

Properties

Prio Recon	Off
Load to viewer	Off
Inline movie	Off
Auto store images	On
Load to stamp segments	On
Load images to graphic segments	Off
Auto open inline display	Off
Wait for user to start	Off
Start measurements	single

Routine

Nr. of slice groups	1
Slices	61
Dist. factor	0 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	230 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
TR	8000 ms
TE	95.0 ms
Averages	6
Concatenations	1
Filter	Prescan Normalize
Coil elements	HE1-4

Contrast

MTC	Off
Magn. preparation	None
Fat suppr.	Fat sat.
Fat sat. mode	Weak
Averaging mode	Long term
Measurements	1
Delay in TR	0 ms
Reconstruction	Magnitude
Multiple series	Off

Resolution

Base resolution	116
Phase resolution	100 %
Phase partial Fourier	6/8
Interpolation	Off
PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	32
Reference scan mode	Separate
Distortion Corr.	Off
Prescan Normalize	On
Normalize	Off
Raw filter	Off
Elliptical filter	Off
Dynamic Field Corr.	Off

Geometry

Nr. of slice groups	1
Slices	61
Dist. factor	0 %
Position	Isocenter
Phase enc. dir.	A >> P
Phase oversampling	0 %
Multi-slice mode	Interleaved
Series	Interleaved
Nr. of sat. regions	0
Position mode	L-P-H
Fat suppr.	Fat sat.
Special sat.	None
Fat sat. mode	Weak
Special sat.	None
Table position	P
Inline Composing	Off

System

Body	Off
HE1	On
HE3	On
NE1	Off
HE2	On
HE4	On
NE2	Off
Position mode	L-P-H
Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
AutoAlign	Head > Brain
Coil Select Mode	Off - AutoCoilSelect
Shim mode	Advanced
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto
? Ref. amplitude 1H	0.000 V
Position	Isocenter
Rotation	0.00 deg
R >> L	230 mm
A >> P	230 mm
F >> H	122 mm
Frequency 1H	123.205716 MHz
Correction factor	1
AddCSaCSatNS 1H	98.851 V
Gain	High
Table position	0 mm
Img. Scale. Cor.	1.000

Physio

1st Signal/Mode	None
Magn. preparation	None
Resp. control	Off

Inline

Inline Composing	Off
Distortion correction	Off

Sequence

Introduction	On
Averaging mode	Long term
Multi-slice mode	Interleaved
Bandwidth	1596 Hz/Px
Optimization	None
Free echo spacing	Off
Echo spacing	0.71 ms
EPI factor	116
RF pulse type	Normal
Gradient mode	Fast
TX/RX delta frequency	0 Hz
TX Nucleus	None
TX delta frequency	0 Hz
Coil elements	HE1-4
Acquisition duration	0 ms

BOLD

Delay in TR	0 ms
Diffusion mode	MDDW
Diff. weightings	1
b-value	0 s/mm ²
Diff. weighted images	On
Trace weighted images	Off
ADC maps	Off
FA maps	Off
Mosaic	Off
Tensor	Off
Distortion Corr.	Off
b-Value >=	0 s/mm ²
Exponential ADC Maps	Off
Invert Gray Scale	Off
Calculated Image	Off

SIEMENS MAGNETOM Skyra syngo MR D13

\\USER\CIND\StudyProtocols\PTSA*dm_ep2d_mono95_64dir_b2000_p2_iso2.0
TA:9:19 PAT:2 Voxel size:2.0x2.0x2.0 mm Rel. SNR:1.00 :epse

Properties

Prio Recon	Off
Load to viewer	Off
Inline movie	Off
Auto store images	On
Load to stamp segments	On
Load images to graphic segments	Off
Auto open inline display	Off
Wait for user to start	Off
Start measurements	single

Routine

Nr. of slice groups	1
Slices	61
Dist. factor	0 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	230 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
TR	8200 ms
TE	97.0 ms
Averages	1
Concatenations	1
Filter	Prescan Normalize
Coil elements	HE1-4

Contrast

MTC	Off
Magn. preparation	None
Fat suppr.	Fat sat.
Fat sat. mode	Weak
Averaging mode	Long term
Measurements	1
Delay in TR	0 ms
Reconstruction	Magnitude
Multiple series	Off

Resolution

Base resolution	116
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off
PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	32
Reference scan mode	Separate
Distortion Corr.	Off
Prescan Normalize	On
Normalize	Off
Raw filter	Off
Elliptical filter	Off
Dynamic Field Corr.	Off

Geometry

Nr. of slice groups	1
Slices	61
Dist. factor	0 %
Position	Isocenter
Phase enc. dir.	A >> P
Phase oversampling	0 %
Multi-slice mode	Interleaved
Series	Interleaved
Nr. of sat. regions	0
Position mode	L-P-H
Fat suppr.	Fat sat.
Special sat.	None
Fat sat. mode	Weak
Special sat.	None
Table position	P
Inline Composing	Off

System

Body	Off
HE1	On
HE3	On
NE1	Off
HE2	On
HE4	On
NE2	Off
Position mode	L-P-H
Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
AutoAlign	Head > Brain
Coil Select Mode	Off - AutoCoilSelect
Shim mode	Advanced
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto
? Ref. amplitude 1H	0.000 V
Position	Isocenter
Rotation	0.00 deg
R >> L	230 mm
A >> P	230 mm
F >> H	122 mm
Frequency 1H	123.205716 MHz
Correction factor	1
AddCSaCSatNS 1H	98.851 V
Gain	High
Table position	0 mm
Img. Scale. Cor.	1.000

Physio

1st Signal/Mode	None
Magn. preparation	None
Resp. control	Off

Inline

Inline Composing	Off
Distortion correction	Off

Sequence

Introduction	On
Averaging mode	Long term
Multi-slice mode	Interleaved
Bandwidth	1596 Hz/Px
Optimization	None
Free echo spacing	Off
Echo spacing	0.71 ms
EPI factor	116
RF pulse type	Normal
Gradient mode	Fast
TX/RX delta frequency	0 Hz
TX Nucleus	None
TX delta frequency	0 Hz
Coil elements	HE1-4
Acquisition duration	0 ms

BOLD

Delay in TR	0 ms
Diffusion mode	MDDW
Diff. weightings	2
b-value 1	0 s/mm ²
Diff. weighted images	On
Trace weighted images	Off
ADC maps	Off
FA maps	On
Mosaic	On
Tensor	Off
Distortion Corr.	Off
b-Value >=	0 s/mm ²
Exponential ADC Maps	Off
Invert Gray Scale	Off
Calculated Image	Off